

ENGEL LABORATORIES

Engel Laboratories · Weiskircher Weg 10 · 63150 Heusenstamm

To whom it may concern

30.07.2021

Individual Sterile Treatment Certificate

Product Description	LC 500 -11 76 x 76 x 122 mm, (L x l x H) , opening 58 mm sterile 500 ml vessels with screw – cap leakproof packed in boxes of 175 pieces
Manufacture batch number	785036
Date of treatment	04 - 2021
Expiration date	04 - 2024

Treatment by electrical source
at minimum dose 10 kGy with quality mark on each case

The products are irradiated by ionizing radiations, in accordance with ANSI / AAMI / ISO111337 to achieve a sterility Assurance Level (SAL) of 10⁻³.

All finished product is tested for performance and bacterial contamination.

This information released by

Engel Laboratories
Quality control for Radient Nord
QF/33AP034