

# ENGEL LABORATORIES

Engel Laboratories · Weiskircher Weg 10 · 63150 Heusenstamm

To whom it may concern

06.01.2025

## Individual Sterile Treatment Certificate

Product Description	LP 291248 175 x 85 x 85 mm, opening 50 mm sterile 1.000 ml vessel with screw - cap leakproof packed in boxes of 72 pieces
Manufacture batch number	D1009A AA                      D 1280 A
Date of treatment	07 – 2024                              10-2024
Expiration date	07 – 2029                              10-2029
Treatment by electrical source at minimum dose	10 kGy with quality mark on each case

The products are irradiated by ionizing radiations, in accordance with ANSI / AAMI / ISO111337 to achieve a sterility Assurance Level (SAL) of 10<sup>-3</sup>.

All finished product is tested for performance and bacterial contamination.

This information released by

Engel Laboratories  
Quality control for Radient Nord  
QF/33AP034