

# ENGEL LABORATORIES

Engel Laboratories · Weiskircher Weg 10 · 63150 Heusenstamm

To whom it may concern

16.10.2024

## Individual Sterile Treatment Certificate

Product Description LC 500-11  
76 x 76 x 122 mm, ( L x l x H ) , opening 58 mm  
sterile 500 ml vessels  
with screw - cap leakproof  
packed in boxes of 175 pieces

Manufacture batch number 902845

Date of treatment 08 - 2024

Expiration date 08 - 2027

Treatment by electrical source  
at minimum dose 10 kGy with quality mark on each case

The products are irradiated by ionizing radiations, in accordance with ANSI / AAMI / ISO111337 to achieve a sterility Assurance Level (SAL) of 10<sup>-3</sup>.

All finished product is tested for performance and bacterial contamination.

This information released by

Engel Laboratories  
Quality control for Radiant Nord  
QF/33AP034