

ENGEL LABORATORIES

Engel Laboratories · Weiskircher Weg 10 · 63150 Heusenstamm

To whom it may concern

19.06.2024

Individual Sterile Treatment Certificate

Product Description LC 500-11
76 x 76 x 122 mm, (L x l x H) , opening 58 mm
sterile 500 ml vessels
with screw - cap leakproof
packed in boxes of 175 pieces

Manufacture batch number 891779

Date of treatment 04 - 2024

Expiration date 04 - 2027

Treatment by electrical source
at minimum dose 10 kGy with quality mark on each case

The products are irradiated by ionizing radiations, in accordance with ANSI / AAMI / ISO111337 to achieve a sterility Assurance Level (SAL) of 10⁻³.

All finished product is tested for performance and bacterial contamination.

This information released by

Engel Laboratories
Quality control for Radiant Nord
QF/33AP034