

ENGEL LABORATORIES

Engel Laboratories · Weiskircher Weg 10 · 63150 Heusenstamm

To whom it may concern

19.04.2022

Individual Sterile Treatment Certificate

Product Description FL 300-04
88 x 75 mm, opening 50 mm
sterile 300 ml vessels
with hinged cap
packed in boxes of 240 pieces

Manufacture batch number 800753

Date of treatment 11 - 2021

Expiration date 11 - 2024

Treatment by electrical source
at minimum dose 10 kGy with quality mark on each case

The products are irradiated by ionizing radiations, in accordance with ANSI / AAMI / ISO111337 to achieve a sterility Assurance Level (SAL) of 10⁻³.

All finished product is tested for performance and bacterial contamination.

This information released by

Engel Laboratories
Quality control for Radient Nord
QF/33AP034