

## Certificate of Sterility

<b>Product</b>	11-702-6110
<b>Lot number</b>	22836
<b>Expiry date</b>	03/2029

nerbe plus GmbH & Co. KG hereby certifies that the product described above has been manufactured in accordance with established manufacturing guidelines and product specifications. The product conforms to all quality requirements. We certify that all mentioned deliveries comply with our valid technical data sheet.

The sterilization process is validated under the standard ISO 11137-1 and controlled under the requirements of the standard of ISO 13485.

nerbe plus certifies products – sterilized by Gamma irradiation or E-beam – to have a sterility assurance level of 10<sup>-6</sup>.

nerbe plus acknowledges that all items, labelled as sterile, product sterility is assured provided that the packaging in which the product is contained is intact and not open or damaged.

22.04.2024  
Quality Management  
**nerbe plus GmbH & Co. KG**

This certificate was issued digitally and is valid without a signature.